## Supplement to Certificate, Issue No. 95117B 10

# **CERTIFICATE OF CALIBRATION**

Issue:- Certificate Number: 95117B
95117B 21 Date of Issue: 15-Jul-21

Approved Signatory: Mark Norfolk

Page 1 of 2 Signed:



#### **Submitter:-**

Mecmesin Limited Newton House Spring Copse Business Park Slinfold West Sussex RH13 0SZ

#### Issued by:-

Kent Scientific Services 8 Abbey Wood Road Kings Hill West Malling Kent ME19 4YT

Tel: 03000 415 100 Fax: 01732 220006

**EQUIPMENT:** Weights Set AH5

SERIAL NUMBER: MB4

MAKE/TYPE: N/A

STANDARDS USED: Set 12412

**DATE RECEIVED:** 25 June 2021

**DATE CALIBRATED:** 29 June 2021

**DETAILS:** 1 Brass

## **MEASUREMENTS:**

Kent Scientific Services method used: CAL-M2, Calibration of Small Masses.

The calibrations took place in a controlled environment with the temperature held between 18°C and 22°C, and with the relative humidity held between 40% and 60%.

The measurement results obtained in the table, where each measured value given represents not the true mass, but the mass of a hypothetical weight of density 8,000 kg.m<sup>-3</sup>, which in air of density 1.2 kg.m<sup>-3</sup> would balance the corresponding weight identified in the first column at 20°C.

The method of weighing was by substitution (Borda's method). In each instance the standard weight used had been calibrated by UKAS Calibration Laboratory number 0474 or 0352 within the previous three years.

The uncertainty of measurements for each of the different denominations is listed in the last column of the table Duplicate weights, where present, are indicated by a dot or dots.

Customer supplied information is notated with a ~, and results relate only to the item(s) calibrated.

Unless otherwise notated, samples are tested in as received condition at Kent Scientific Services.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Certificate No.: 95117B Page 2 of 2

# TABLE OF MEASUREMENT RESULTS

Nominal	Measured	Error from	Estimated
Mass	Value	Nominal	Uncertainty
0.5 N	50.956 29 g	- 4.53 mg	± 1.61 mg

The basis for conversion between force units and mass units is that a 1kg mass will experience a force of g newtons where g is the strength of the local gravitational field. At Kent Scientific Services the estimated local g =  $9.81146 \text{ms}^{-2}$ .

**END OF RESULTS**