

Internal Calibration Equipment Certificates Index

Revision: 24th October 2022

WEIGHTS & MASSES

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due	Serial Numbers	Details
95511_21	30-12-21	Pass	DECEMBER 2023	1957 to 1964, 2001 to 2062, 3020 to 3026, 3362A to 3362M, 3500 to 3504 3506 to 3521 & TB4	SET MC3 111 x Cast Iron
91986_21	28-12-2017	Pass	DECEMBER 2022	1 to 25, 2327 & 2328	SET MS2 26 x Stainless Steel
95510	30-12-2021	Pass	DECEMBER 2023	1927,1929, 1931-1933, 1935, 1966-1969, 1970, 1980, 1981, 1926, 1930, 1971- 1972, 1925, 1973- 1974,1937	SET MC2 21 x Cast Iron
95512	31-12-2021	Pass	DECEMBER 2023	3034	SET MS1 13 x Stainless Steel
95509	30-12-2021	Pass	DECEMBER 2023	1910, 1911, 1913-1916, 1921, 1928, 1975-1976.	SET MC1 10 x Cast Iron
94491	19-08-2020	Pass	SEPTEMBER 2023	CL1 to CL5	SET MB3 5 Brass
92010	05-01-2018	Pass	JANUARY 2023	S1 01 to S1 28	SET S1
93334_21	16-05-2019	Pass	MAY 2024	S2 01 to S2 27	SET S2
93335_21	20-05-2019	Pass	MAY 2024	S3 01 to S3 27	SET S3
92622	10-08-2018	Pass	AUGUST 2023	S4 01 to S4 14	SET S4 - A
92623	10-08-2018	Pass	AUGUST 2023	S4 15 to S4 18	SET S4 - B
93835	27-11-2019	Pass	NOVEMBER 2023	S5 01 to S5 27	SET S5

OTHER EQUIPMENT

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due	Serial Numbers	Details
95764	11-05-2022	Pass	MAY 2023	4101220/1053	TM0260 Hanna Checktemp 1 Temperature Indicator
020050	18-08-2022	Pass	AUGUST 2025	SGS0591	TORQUE CALIBRATION BEAM
018079	21-05-2021	Pass	MAY 2024	SGS0491	TORQUE CALIBRATION BEAM
020328	18-10-2022	Pass	OCTOBER 2025	SGS0593	TORQUE CALIBRATION BEAM
2208006 2208007	05-08-2022 08-08-2022	Pass	AUGUST 2024	502318 (TM0295) 502318 (TM0294)	50kN Standard 10kN Standard

N.B. Weights, masses and other equipment as listed above placed within the calibration laboratory must temperature acclimatise for at least two hours before use.