Supplement to Certificate, Issue No. 96835_10							
CERTIFICATE OF CALIBRATION							
Issue:- Certificate Number:	96835						
96835_21 Date of Issue:	11-Oct-23						
Approved Signatory:	Mark Norfolk	<b>UKAS</b> CALIBRATION					
Page 1 of 2 Signed:	here	0352					
Submitter:- Mecmesin Limited Newton House Spring Copse Business Park Slinfold West Sussex RH13 0SZ			Issued by:- Kent Scientific Services 8 Abbey Wood Road Kings Hill West Malling Kent ME19 4YT Tel: 03000 415 100 Fax: 01732 220006				
EQUIPMENT:	Weights						
SERIAL NUMBER:	MB3						
MAKE/TYPE:	N/A						
STANDARDS USED:	Set 12412						
DATE RECEIVED:	29 August 2023						
DATE CALIBRATED:	20 September 2023	3					
DETAILS:	5 Brass						

## **MEASUREMENTS:**

Kent Scientific Services method used: CAL SMALL, Calibration of Small Masses.

The calibrations took place in a controlled environment with the temperature held between 18°C and 22°C, and with the relative humidity held between 40% and 60%.

The measurement results obtained in the table, where each measured value given represents not the true mass, but the mass of a hypothetical weight of density  $8,000 \text{ kg.m}^{-3}$ , which in air of density  $1.2 \text{ kg.m}^{-3}$  would balance the corresponding weight identified in the first column at  $20^{\circ}$ C.

The method of weighing was by substitution (Borda's method). In each instance the standard weight used had been calibrated by UKAS Calibration Laboratory number 0474, 0260 or 0352 within the previous three years. The uncertainty of measurements for each of the different denominations is listed in the last column of the table Duplicate weights, where present, are indicated by a dot or dots.

Customer supplied information is notated with a  $\sim$ , and results relate only to the item(s) calibrated. Unless otherwise notated, samples are tested in as received condition at Kent Scientific Services.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Certificate No.: 96835

Page 2 of 2

## **TABLE OF MEASUREMENT RESULTS**

Identity Mark	Nominal Mass	Measured Value	Error from Nominal	Estimated Uncertainty
CL1	0.5 N	50.948 09 g	- 12.72 mg	± 0.61 mg
CL2	0.5 N	50.949 97 g	- 10.85 mg	± 0.61 mg
CL3	0.5 N	50.950 12 g	- 10.69 mg	± 0.61 mg
CL4	0.5 N	50.949 92 g	- 10.90 mg	± 0.61 mg
CL5	0.5 N	50.947 26 g	- 13.55 mg	± 0.61 mg

The basis for conversion between force units and mass units is that a 1kg mass will experience a force of g newtons where g is the strength of the local gravitational field. t Kent Scientific Services the estimated local  $g = 9.81146 \text{ ms}^{-2}$ . #

# Updated with force conversion information

END OF RESULTS

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.