CERTIFICATE OF CALIBRATION

ISSUED BY MP CALIBRATION SERVICES

DATE OF ISSUE: 14 December 2023

CERTIFICATE NUMBER: 397349



0228

PAGE 1 OF 1 PAGE (S)

NAME: T Rutter

عاليه عيني المنطوع التراس

SIGNED:

APPROVED SIGNATORY

43 HAVILAND RD FERNDOWN INDUSTRIAL ESTATE WIMBORNE DORSET **BH21 7RY**

Telephone:

01202 624468

Fax:

01202 625132

SUBMITTED BY:

DATE OF CALIBRATION: 14/12/2023

CUSTOMER:

PPT Group UK Ltd (Slinfold) t/a Mecmesin Ltd Slinfold, Stane Lane, West Sussex, Horsham, RH13 0SZ

DESCRIPTION:

Micrometer Setting Rod(s) Metric

SIZE(S): 200 mm

MANUFACTURER: Mitutoyo

SERIAL NO:

LB₂

BASIS OF TEST:

BS 870: 2008 / Manufacturers Specification

TEMPERATURE: 20 °C ± 2 °C

RESULTS OF TEST:

The accuracy of the setting rod(s) was checked using the laboratories master gauge blocks, length bars and electronic comparator. The mean measured results obtained are detailed below:

Nom.Size

Tolerance (+/-)

Result

200 mm

0.003 mm

200.0003 mm

Uncertainty of measurement

 \pm 0.001 mm + (0.008mm x Length in Metres)

Out of Specification readings will be indicated by *

End of report This report relates only to the item with the serial number shown above

Decision rule: Unless otherwise requested MP Calibration Services Ltd shall issue a conformity statement based on a binary decision rule (ie. simple acceptance of pass/fail) where the acceptance zone is equal to the tolerance zone stated in the standard used as the basis of test, and a measurement uncertainty which is not greater than the tolerance. Where measurement uncertainty is greater than the tolerance zone it is not possible to state conformity

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2,providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.