

On-site Equipment Calibration Certificates Index

Revision: 12-02-2025

Weights and Masses

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due Date	Serial Numbers	Details
97643	08-10-2024	Pass	October 2025	S2E 1N, M02-M14 (no M04)	SET AH1 14 x Cast Iron 1 x Brass
97739	15-11-2024	Pass	November 2025	X	SET AH2 10 x Brass
97738-21	15-11-2024	Pass	November 2025	S2A to S2N, PO15, MB2	SET AH3 14 x Cast Iron, 1 x Brass
97863	15-01-2025	Pass	January 2026	P	SET AH6 10 x Brass
97438	07-06-2024	Pass	June 2025	PO1 to PO14, S2E	SET AH8 14 x Cast iron 1 x Brass
97660	10-10-2024	Pass	October 2025	T	SET AH9 x 10 Brass
96505_21	28-03-2024	Pass	March 2025	JO1 to J14, Z	SET FR1 x 14 Cast iron & 10 Brass

Length Bars

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due Date	Serial Numbers	Details
416782	23-08-2024	Pass	23-08-2025	TM0112	Length Bar 200mm
416783	23-08-2024	Pass	23-08-2025	TM0113	Length Bar 300mm
416784	23-08-2024	Pass	23-08-2025	TM0114	Length Bar 400mm
416780	23-08-2024	Pass	23-08-2025	TM0115	Length Bar 100mm
				TM0116	Length Bar 100mm
				TM0117	Length Bar 200mm
				TM0118	Length Bar 300mm
420861	22-10-2024	Pass	22-10-2025	TM0119	Length bar 400mm
416785	23-08-2024	Pass	23-08-2025	TM0120	Length Bar 600mm
				LB12	Length Bar 400mm
				TM0127	Length Bar 200mm
420866	22-10-2024	Pass	22-10-2025	TM0287	Length Bar 600mm
				TM0129	Length Bar

On-site Equipment Calibration Certificates Index

Revision: 12-02-2025

					400mm
				TM0127	Length Bar 200mm
399378	18-01-2024	Pass	17-01-2026	TM0280	Length bar 25mm
420862	22-10-2024	Pass	22-10-2025	TM0252	Length bar 100mm
420864	22-10-2024	Pass	22-10-2025	TM0253	Length bar 200mm
420865	22-10-2024	Pass	22-10-2025	TM0254	Length bar 300mm
372965	18-01-2024	Pass	17-01-2026	TM0281	Length bar 50mm
372966	18-01-2024	Pass	17-01-2026	TM0282	Length bar 75mm
426852	15-01-2024	Pass	15-01-2027	TM0313	Length bar 25mm
426853	15-01-2024	Pass	15-01-2027	TM0314	Length bar 50mm
426855	15-01-2024	Pass	15-01-2027	TM0315	Length bar 75mm
				TM0326	Length Bar 600mm
414128	22-07-2024	Pass	22-07-2025	TM0453	Length bar 25mm
414132	22-07-2024	Pass	22-07-2025	TM0454	Length bar 50mm
414133	22-07-2024	Pass	22-07-2025	TM0455	Length bar 75mm
419206	27/09/2024	Pass	27/09/2025	TM0484	Length bar 25mm
419208	27/09/2024	Pass	27/09/2025	TM0485	Length bar 50mm
419209	27/09/2024	Pass	27/09/2025	TM0486	Length bar 75mm
419220	27/09/2024	Pass	27/09/2025	TM0487	Length Bar 100mm
419221	27/09/2024	Pass	27/09/2025	TM0488	Length Bar 200mm
419223	27/09/2024	Pass	27/09/2025	TM0489	Length Bar 300mm
419224	27/09/2024	Pass	27/09/2025	TM0490	Length Bar 400mm
420887	22-10-2024	Pass	22-10-2025	TM0491	Length bar 600mm

On-site Equipment Calibration Certificates Index

Revision: 12-02-2025

Transfer Standard Load Cells

Certificate/Serial Number	Calibration Date	Pass Fail	Next Calibration Due Date	Serial Numbers	Details
2404025	08-04-2024	Pass	April 2026	TM0198	5kN
2404026	09-04-2024	Pass	April 2026	TM0199	25kN
2310025	11-10-2023	Pass	October 2025	TM0300	5kN
2310026	12-10-2023	Pass	October 2025	TM0301	25kN
2401021	12-01-2024	Pass	January 2026	TM0417	5kN
2401022	12-01-2024	Pass	January 2026	TM0419	25kN