

Internal Calibration Equipment Certificates Index

Revision: 22-10-2025

WEIGHTS & MASSES

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due	Serial Numbers	Details
97110	04-01-2024	Pass	JANUARY 2026	1957 to 1964, 2001 to 2062, 3020 to 3026, 3362A to 3362M, 3500 to 3504 3506 to 3521 & TB4	SET MC3 112 x Cast Iron
96291	30-12-22	Pass	DECEMBER 2027	1 to 25, 2327 & 2328	SET MS2 26 x Stainless Steel
97109	22-12-2023	Pass	DECEMBER 2025	1927,1929, 1931-1933, 1935, 1966-1969, 1970, 1980, 1981, 1926, 1930, 1971-1972, 1925, 1973-1974,1937	SET MC2 21 x Cast Iron
97107	21-12-2023	Pass	DECEMBER 2025	3034	SET MS1 13 x Stainless Steel
97108 / 97141	22-12-2023 / 25-01-2024	Pass	DECEMBER 2025	1875, 1912, 1915, 1916, 1921,1928, 1936, 1939, 1942, 1976.	SET MC1 10 x Cast Iron
96835	20-09-2023	Pass	SEPTEMBER 2026	CL1 to CL5	SET MB3 5 Brass
96340	23-01-2023	Pass	JANUARY 2028	S1 01 to S1 28	SET S1
97350	26-04-2024	Pass	APRIL 2029	S2 01 to S2 27	SET S2
97429	04-06-2024	Pass	JUNE 2029	S3 01 to S3 27	SET S3
96802	20-09-2023	Pass	SEPTEMBER 2028	S4 01 to S4 14	SET S4 - A
96803	22-09-2023	Pass	SEPTEMBER 2028	S4 15 to S4 18	SET S4 - B
97781	04-12-2024	Pass	DECEMBER 2029	S5 01 to S5 27	SET S5

OTHER EQUIPMENT

Certificate Number	Calibration Date	Pass Fail	Next Calibration Due	Serial Numbers	Details
98176	18-06-2025	Pass	June 2026	4101220/1053	TM0260 Hanna Checktemp 1 Temperature Indicator
024974	03-07-2025	Pass	JULY 2028	SGS0591	TORQUE CALIBRATION BEAM
023009	07-05-2024	Pass	MAY 2027	SGS0584	TORQUE CALIBRATION BEAM
025519	16-10-2025	Pass	OCTOBER 2028	SGS0593	TORQUE CALIBRATION BEAM
2407041 2407042	29-07-2024 26-07-2024	Pass	JULY 2026	502318 (TM0295) 502318 (TM0294)	50kN Standard 10kN Standard

N.B. Weights, masses and other equipment as listed above placed within the calibration laboratory must temperature acclimatise for at least two hours before use.