

CERTIFICATE OF CALIBRATION

ISSUED BY: Torus Measurement Systems Ltd.

DATE OF ISSUE: 16th October 2025 SERIAL No: 025519



0773



Torus Measurement Systems Ltd.

Nedge Hill Science Park, Telford

Shropshire TF3 3AJ

United Kingdom

Tel: +44 (0) 1952 210020

Fax: +44 (0) 1952 299804

Email: tms.inspection@torus-group.com

A Member of the TORUS GROUP

www.torus-group.com

PAGE 1 of 3 PAGES

APPROVED SIGNATORIES

D. Ball

E. Warlow ✓

P. Wood

CUSTOMER: Mecmesin (PPT Group)
Newton House, Spring Copse Business Park, Stane St., Slinfold, RH13 0SZ

ORDER NO: PO140631-1

DESCRIPTION: Torque Beam

RANGE/SIZE: See Pages 2 - 3

MANUFACTURER: N/A

QUANTITY: 1

SERIAL NO: SGS0593

TECHNICAL STANDARD: N/A

DRAWING NO: PDV13110-17 ISS 02

DATE RECEIVED: 6th October 2025

REPORT:

This Jig/ Fixture has been examined and measured at 20°C ±1°C using a Global CMM, a Reference Sphere and a Digital Thermometer & Probe. The mean of the measured results obtained are stated on the following page/s of this report.

* Denotes 'Out of Tolerance' to the above drawing

x Denotes errors before Repair/ Resetting

Repairs/ Resetting prior to calibration -/ NO

CALIBRATION EQUIPMENT: Lab 51, Lab 136 & Lab 106

DATE CALIBRATED: 16th October 2025

Signature:

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Simple acceptance shared risk (Ref JCGM 106:2012, section 8.2) where the expanded uncertainty for a coverage factor of $K=2$ associated with the result/estimate is $\leq$ the tolerance.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: Torus Measurement Systems Ltd

UKAS Accreditation Calibration Laboratory No.0773

DATE OF ISSUE

16th October 2025

SERIAL NUMBER

025519

PAGE 2 of 3 PAGES

CUSTOMER: Mecmesin (PPT Group)

DESCRIPTION: Torque Beam

DRAWING: PDV13110-17 ISS 02

SERIAL NUMBER: SGS0593

DESCRIPTION	AXIS	REQUIREMENT	MEASURED	DEVIATION	UNCERTAINTY UNIT
CIRCLE LEFT 1	Y	100.000 +0.100/-0.100	99.990	-0.010	± 0.0057 mm
CIRCLE LEFT 2	Y	95.000 +0.100/-0.100	94.990	-0.010	± 0.0057 mm
CIRCLE LEFT 3	Y	90.000 +0.100/-0.100	89.987	-0.013	± 0.0057 mm
CIRCLE LEFT 4	Y	85.000 +0.100/-0.100	84.987	-0.013	± 0.0057 mm
CIRCLE LEFT 5	Y	80.000 +0.100/-0.100	79.986	-0.014	± 0.0057 mm
CIRCLE LEFT 6	Y	75.000 +0.100/-0.100	74.985	-0.015	± 0.0057 mm
CIRCLE LEFT 7	Y	70.000 +0.100/-0.100	69.986	-0.014	± 0.0057 mm
CIRCLE LEFT 8	Y	65.000 +0.100/-0.100	64.987	-0.013	± 0.0057 mm
CIRCLE LEFT 9	Y	60.000 +0.100/-0.100	59.985	-0.015	± 0.0057 mm
CIRCLE LEFT 10	Y	55.000 +0.100/-0.100	54.985	-0.015	± 0.0057 mm
CIRCLE LEFT 11	Y	50.000 +0.100/-0.100	49.985	-0.015	± 0.0057 mm
CIRCLE LEFT 12	Y	45.000 +0.100/-0.100	44.988	-0.012	± 0.0057 mm
CIRCLE LEFT 13	Y	40.000 +0.100/-0.100	39.989	-0.011	± 0.0057 mm
CIRCLE LEFT 14	Y	35.000 +0.100/-0.100	34.987	-0.013	± 0.0057 mm
CIRCLE LEFT 15	Y	30.000 +0.100/-0.100	29.988	-0.012	± 0.0057 mm
CIRCLE LEFT 16	Y	25.000 +0.100/-0.100	24.986	-0.014	± 0.0057 mm

NOTE:

All above calibrated in metric

Measured value is the average of 3 runs

Measured values only relate to the equipment identified above

Side marked L is LEFT, unmarked side is RIGHT

Signature:



CERTIFICATE OF CALIBRATION

ISSUED BY: Torus Measurement Systems Ltd

UKAS Accreditation Calibration Laboratory No.0773

DATE OF ISSUE

16th October 2025

SERIAL NUMBER

025519

PAGE 3 of 3 PAGES

CUSTOMER: Mecmesin (PPT Group)

DESCRIPTION: Torque Beam

DRAWING: PDV13110-17 ISS 02

SERIAL NUMBER: SGS0593

DESCRIPTION	AXIS	REQUIREMENT	MEASURED	DEVIATION	UNCERTAINTY UNIT
CIRCLE RIGHT 1	Y	-100.000 +0.100/-0.100	-100.019	-0.019	± 0.0057 mm
CIRCLE RIGHT 2	Y	-95.000 +0.100/-0.100	-95.019	-0.019	± 0.0057 mm
CIRCLE RIGHT 3	Y	-90.000 +0.100/-0.100	-90.018	-0.018	± 0.0057 mm
CIRCLE RIGHT 4	Y	-85.000 +0.100/-0.100	-85.018	-0.018	± 0.0057 mm
CIRCLE RIGHT 5	Y	-80.000 +0.100/-0.100	-80.018	-0.018	± 0.0057 mm
CIRCLE RIGHT 6	Y	-75.000 +0.100/-0.100	-75.017	-0.017	± 0.0057 mm
CIRCLE RIGHT 7	Y	-70.000 +0.100/-0.100	-70.016	-0.016	± 0.0057 mm
CIRCLE RIGHT 8	Y	-65.000 +0.100/-0.100	-65.014	-0.014	± 0.0057 mm
CIRCLE RIGHT 9	Y	-60.000 +0.100/-0.100	-60.014	-0.014	± 0.0057 mm
CIRCLE RIGHT 10	Y	-55.000 +0.100/-0.100	-55.013	-0.013	± 0.0057 mm
CIRCLE RIGHT 11	Y	-50.000 +0.100/-0.100	-50.013	-0.013	± 0.0057 mm
CIRCLE RIGHT 12	Y	-45.000 +0.100/-0.100	-45.014	-0.014	± 0.0057 mm
CIRCLE RIGHT 13	Y	-40.000 +0.100/-0.100	-40.012	-0.012	± 0.0057 mm
CIRCLE RIGHT 14	Y	-35.000 +0.100/-0.100	-35.011	-0.011	± 0.0057 mm
CIRCLE RIGHT 15	Y	-30.000 +0.100/-0.100	-30.008	-0.008	± 0.0057 mm
CIRCLE RIGHT 16	Y	-25.000 +0.100/-0.100	-25.013	-0.013	± 0.0057 mm

NOTE:

All above calibrated in metric

Measured value is the average of 3 runs

Measured values only relate to the equipment identified above

Side marked L is LEFT, unmarked side is RIGHT

End of Report

Signature: 