

## CERTIFICATE OF VERIFICATION

**Date of issue:** **17/12/2025**
**Certificate number:** **TM0373**
**Customer reference:** **Mecmesin Calibration**
**Original**
**Issued to:**
**Newton House**  
**Spring Copse Business Park**  
**Slinfold**  
**West Sussex**  
**RH13 0SZ**
**Instrument type:** **MT 2.5dV & Linear**
**Encoder**
**Serial number(s):** **TM0373, (1448666)**
**Date of verification:** **17/12/2025**

The calibration was performed using the following method and equipment.

**Verification method:** **Mecmesin Internal (Speed)** details available on request
**Identification of standards used:** **LB1A LB2A LB3A LB4A**
**Approved signature**

**Name printed here**
**T Dowell**
**(Authorised Technician)**

This is to certify that the operational speed of the above instrument was verified using the method stated above.

 Where possible, this instrument was also independently checked using a tachometer.  
 'PASS / FAIL' is determined by the specification stated in the product datasheet.

 This certificate may not be reproduced other than in full, except with prior written approval of the issuing laboratory. If, upon reproduction, only part of this report is copied, Mecmesin will not bear any responsibility for content, purport and conclusions of that reproduction. This report has legal value only when printed on Mecmesin paper and furnished with an authorised signature. Digital versions of this report have no legal value. The terms and conditions of Mecmesin Ltd (to be found at [www.mecmesin.com](http://www.mecmesin.com)) are applicable on all services provided by Mecmesin Ltd.

<b>Nominal Length</b> mm	<b>Instrument reading</b> mm	<b>PASS / FAIL</b>
0	0	PASS
100	99.990	PASS
200	199.995	PASS
300	300.010	PASS
400	400.025	PASS

### End of Report

## Technical Advisory Note

### – Internal Verification of TM0373

TM0373 is the reference used for Mecmesin's internal Long Travel Extensometer Calibration system. This system has been specifically designed for the calibration of long travel extensometers (LTE). The calibration system is comprised of a Mecmesin motorised force tester and an accompanying *Mitutoyo* linear encoder (s/n 1448666).


Mechanically, the linear encoder is directly attached to the force tester's moving crosshead. Electronically, the encoder is connected to the control board of the force tester. Consequently, the displacement displayed by the motorised force tester is derived from appropriate conditioning of the encoder's quadrature output signal.

The equipment under test (Long Travel Extensometer) can be fitted to the force tester. The relative extension of the LTE is determined by locating upper arm of the LTE to the moving crosshead and fixing the lower arm in place. Both the linear encoder and the upper LTE arm are attached to crosshead. Therefore, the conditioned displacement output of the LTE can be calibrated against the displacement output of the linear encoder beam whilst the Mecmesin force tester moves its crosshead upwards.

The accuracy of the *Mitutoyo* linear encoder's displacement output, once fitted to the Mecmesin force tester, is regularly checked by internally calibrating the encoder against micrometer setting rods. The method involves using a touch-on load to establish a datum point. A setting rod is then placed in between the datum point and the force tester crosshead, and a touch-on load is used once again to establish the location of the setting rod.

The accuracy of the internal calibration of the *Mitutoyo* linear encoder and any sources of uncertainty of the internal calibration have been included in the overall uncertainty budget for the Long Travel Extensometer Calibration System.

The micrometer settings rods are regularly calibrated by a UKAS accredited laboratory. Therefore, the traceability chain to National standards is maintained through the use of these micrometer setting rods as described above. If the linear encoder is damaged in use, the encoder will be inspected and verified before being used again.



Dr Patrick Collins  
Technical Director  
Tel: +44 (0)1403 799979  
M: +44 (0)7939 214059  
Fax: +44 (0)1403 799975  
Email: [pat.collins@mecmesin.com](mailto:pat.collins@mecmesin.com)

### Force & Torque Test Solutions

Mecmesin Ltd, Newton House, Spring Copse Business Park, Slinfold, West Sussex, RH13 0SZ, United Kingdom  
t: +44 (0) 1403 799979 • f: +44 (0) 1403 799975 • UK Sales: +44 (0) 1403 799910 • Export Sales: +44 (0) 1403 799916  
Service & Calibration: t: +44 (0) 1403 799920 • e: [info@mecmesin.com](mailto:info@mecmesin.com) • [www.mecmesin.com](http://www.mecmesin.com)