



Issued by
Trescal Ltd
Park Gate Close, Bredbury Park Way
Bredbury, Stockport
SK6 2SL, UK
Tel: +44 (0)161 4067878

APPROVED SIGNATORY

Hazem Salama (892)
(Signed electronically)

Customer:
PPT Group UK Ltd t/a Mecmesin, Newton House, Spring Copse Business Park
Slindfold, Horsham, West Sussex, RH13 0SZ

Location of calibration:
Trescal Ltd, Park Gate Close, Bredbury Park Way
Bredbury, Stockport, SK6 2SL, UK

Equipment Details:

Description:	Speed Verification Tool		
Manufacturer:	ELV	Customer Ref:	TM0435
Type No:	LSU 100	Date of Receipt:	10 Oct 2025
		Order No:	PO140770-1
Serial No:	SEQ2878979	Our Reference:	76026783
Calibrated By:	Alexandre Duarte	Date of Calibration:	16 Oct 2025

Calibration Summary

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.
Assessment of conformance has been undertaken in accordance with the agreed decision rule detailed within this certificate.

Status on Receipt: Measured Values Only

No assessment of conformance has been undertaken.

Status on Despatch: Measured Values Only

No assessment of conformance has been undertaken.

Action(s) Taken: Full Calibration

Ambient Conditions: Temperature: 20 ± 2 °C

Humidity: 35 ± 20 %rh

Customer requested calibration due date:

The results given within this certificate only relate to the item calibrated. The expanded uncertainties quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Agreed and applied Decision Rule used to determine conformance:

It has been stipulated by the customer that assessment of conformance is not required. (Trescal Rule 0)
Reference: Trescal document TRE/DEC/0004 Issue 1

Calibration Equipment Used

Asset No	Due Date	Asset No	Due Date
120518	30/11/2026	E492	30/09/2026
FC160	31/10/2025	FC332a	28/02/2026
FC406	31/10/2025	FC552	31/01/2026
FC564	31/01/2026	FC571	01/04/2026

The following information and any identified symbols are used by Trescal in determining the conformance of the device at each measurement point, which is then evaluated in line with the requested decision rule. Where this decision rule is of a "Binary Rule" type then the "Conditional Pass & Fail" statuses are treated as specified by the chosen decision rule. The overall conformance statement determined is given on page one of this certificate.

Key to results annotations, where applicable


- * The measurements marked thus have been determined to be outside the specification, due allowance having been made for the expanded uncertainty.
- ! The measurements marked thus are within the specification by a margin less than the expanded uncertainty.
- ^ The measurements marked thus are outside the specification by a margin less than the expanded uncertainty.
- @ The expanded uncertainty is significant in relation to the unit specification and it is therefore not possible to demonstrate conformance unambiguously.
- \$ The measurements marked thus do not lie within the scope of the Laboratory's prevailing UKAS accreditation but are reported herein for completeness.

Basis of Test: Measured results.

This instrument has been measured using laboratory standards for the distance between the laser sensors operating the stopwatch (Figure 1), with the datum laser being at the bottom of the column. Repeat readings were taken to determine the accuracy and the measured results reported in the following tables:

Position reference	Nominal length (mm)	Start position (mm)	Stop position (mm)	Measured length (mm)
1	21	68.105	88.912	20.807
2	21	68.103	88.923	20.82
3	21	68.111	88.917	20.806
4	21	68.123	88.899	20.776
5	21	68.118	88.905	20.787
Mean Value (mm)		68.112	88.911	20.799
Maximum variation (mm)		0.020	0.024	0.044

Position reference	Nominal length (mm)	Start position (mm)	Stop position (mm)	Measured length (mm)
1	100	68.105	168.473	100.368
2	100	68.103	168.488	100.385
3	100	68.111	168.455	100.344
4	100	68.123	168.445	100.322
5	100	68.118	168.483	100.365
Mean value (mm)		68.112	168.469	100.357
Maximum variation (mm)		0.020	0.043	0.063



The dovetail bracket (Figure 2 & 3) was measured for the distance between faces with the following results:

Feature	Position	Nominal Distance (mm)	Measured Results (mm)	Mean Value (mm)	Parallelism (mm)
Upwards Distance	1	20.0	20.044	20.036	0.018
	2		20.039		
	3		20.036		
	4		20.026		
	5		20.036		

Feature	Position	Nominal Distance (mm)	Measured Results (mm)	Mean Value (mm)	Parallelism (mm)
Downwards Distance	1	20.0	19.955	19.983	0.038
	2		19.992		
	3		19.993		
	4		19.993		
	5		19.983		

Figure 2





Figure 3



Uncertainty of measurement : ± 0.003 mm (Linear)
 ± 0.005 mm (Laser displacement)

Calibration Procedure : QCD/CALP/18
 Our Reference : AFD/482180

Calibration Procedure: The instrument was placed in the laboratory for 24 hours prior to calibration for stabilisation purposes. Tests were made by comparison with a standard counter at the given indications. The equipment was calibrated in a controlled environment using devices having known and traceable values. The uncertainties reported refer to the measured values only with no account being taken of the instrument ability to maintain its calibration.

Time Checks

<u>UUT Indication</u>	<u>Standard Indication</u>	<u>Standard Equivalent</u>
00 hr 00 min 05.214 sec	5.159 sec	0 min 05.159 sec
00 hr 00 min 10.309 sec	10.248 sec	0 min 10.248 sec
00 hr 00 min 30.125 sec	30.101 sec	0 min 30.101 sec
00 hr 01 min 00.406 sec	60.360 sec	1 min 00.360 sec
00 hr 05 min 00.295 sec	300.176 sec	5 min 00.176 sec
00 hr 10 min 00.244 sec	600.152 sec	10 min 00.152 sec

The overall uncertainty in the measurement was:

$\pm (0.06)$ seconds of indication

Specification taken from:

Measured results only no specifications available.

End of Results