



TM0447

Calibration Certificate



3590VHR Calibration stand

Doc #: 5464.022726.1200

Model: 3590VHR

XL-80 Laser, Calibrated

20-Nov-23

Model S/N: A5464

S/N: 5FM252

Certificate Number: 5FM252-231120-00

Calibration Date: 27-Feb-2026

Traceability Info

Ambient Temperature: (°F) 71.9

Humidity: 27%

Scale SN: 2UMU77

Model #	Certificate No.	Cal Date	Cal Lab
RUK27030	2022080011-LL03	6-Aug-22	NPL
MTE/A163	23-70302	7-Jun-23	Nationwide
XL80 REF5	86H576-231006-00	6-Oct-23	Renishaw

As Found / As Left (PASS, No Change)

For use in extensometer calibrations in accordance with ASTM E83, ISO 9513, or any other standards; the requirements of the standard should be used to determine the class to which any particular extensometer may be calibrated. This calibration stand was calibrated using a method developed at Epsilon Technology and detailed in the Calibration and Maintenance Procedure. The information on this certificate applies only to the item with the serial number listed above. Reference displacement and environmental conditions were measured using the laser interferometer listed on this certificate. The reference laser is routinely calibrated by an accredited laboratory, with traceability as indicated above.

Verified direction of travel is upward. Uncertainty of Calibration: U = 5.5µin, increasing to 16.9µin at 2.00in; coverage factor k=2, approximate confidence level 95%. Results below are raw data and are not compensated for temperature or uncertainty. It is the responsibility of the end user to determine if it is appropriate for your specific application. If more than one calibration certificate exists, the most recent should be considered to supercede all previous certificates.

Full Range, 0-2.0 (in)

Laser (in)	3590VHR (in)	Error (in)
0.000000	0.000000	0.000000
0.200016	0.200013	-0.000003
0.400062	0.400057	-0.000005
0.600024	0.600024	-0.000001
0.800005	0.799995	-0.000010
1.000027	1.000002	-0.000026
1.200024	1.199988	-0.000036
1.400028	1.399998	-0.000031
1.600024	1.600001	-0.000023
1.800021	1.799993	-0.000027
2.000009	1.999980	-0.000029
1.000029	1.000008	-0.000021
0.000011	0.000042	0.000030

Small Increments, from ~1.0 (in)

Laser (in)	3590VHR (in)	Error (in)
0.000000	0.000000	0.000000
0.000981	0.000981	0.000000
0.001981	0.001981	0.000000
0.002982	0.002981	-0.000001
0.004974	0.004973	-0.000001
0.009974	0.009969	-0.000005
0.014987	0.014980	-0.000007
0.020028	0.020021	-0.000007

Small Increments, from ~2.0 (in)

Laser (in)	3590VHR (in)	Error (in)
0.000000	0.000000	0.000000
0.001010	0.001011	0.000000
0.001989	0.001989	0.000000
0.003006	0.003005	-0.000001
0.005026	0.005024	-0.000001
0.010038	0.010036	-0.000002
0.015002	0.014998	-0.000004
0.019992	0.019990	-0.000002

Small Increments, from ~0.0 (in)

Laser (in)	3590VHR (in)	Error (in)
0.000000	0.000000	0.000000
0.001015	0.001013	-0.000001
0.002007	0.002006	-0.000001
0.003005	0.003003	-0.000001
0.005010	0.005008	-0.000001
0.010012	0.010012	0.000000
0.015003	0.015003	0.000001
0.020014	0.020013	-0.000001

Meets manufacturer's accuracy specifications and ISO 9513 Class 0,5 and ASTM E83 Class B1 for gage lengths down to 6mm

Remarks:

Calibrated by: Rachel Meepore
Epsilon Technology Corp.
3975 S. Hwy. 89, Jackson, WY 83001, USA

Date: 27 Feb 2026 Issued by: JS
PHONE: (307) 733-8360
FAX: (307) 733-8375

Date: March 2, 2026

