

CERTIFICATE OF CALIBRATION

Issued By: Mitutoyo (UK) Ltd. Calibration Laboratory

Date of Issue: 05-May-2026 Certificate No.: 340744



0332



Calibration Laboratory:

Mitutoyo (UK) Ltd
6 Banner Park, Wickmans Drive
Coventry, West Midlands
CV4 9XA, United Kingdom
T +44 (0)2476 426300
F +44 (0)2476 426339
calibration@mitutoyo.co.uk

Head Office:

Mitutoyo (UK) Ltd
West point Business Park, Joule Road
Andover, Hampshire
SP10 3UX, United Kingdom
T +44 (0)1264 353123
F +44 (0)1264 354883
enquiries@mitutoyo.co.uk

Page 1 of 2

Approved Signatory:

G. Adams

Customer:

PPT Group UK Ltd (Slinfold)
t/a Mecmesin Limited
Slinfold

Manufacturer

Mitutoyo

Description

50 mm Setting Rod

Identification

167-102

Serial No.

4111846

TM 0507

Calibration Conditions

Ambient Temperature
20.0 +/- 0.5 Degrees C.

Basis of Calibration

To calibrate at the above conditions using traceable calibrated results to the Manufacturer's Specification.

Date of Calibration

05-May-2026

Method of Calibration

This item has been calibrated to the above standard using traceable calibrated equipment listed on this certificate.

Estimated uncertainty of measurement:

$\pm 1.0 + (5 \times \text{length in metres})$ micrometres.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Authorised By

CERTIFICATE OF CALIBRATION



UKAS Accredited Calibration Laboratory No. 0332

Mitutoyo (UK) Ltd. 6 Banner Park, Wickmans Drive
Coventry, West Midlands CV4 9XA, United Kingdom

Certificate No.: 340744

Page 2 of 2

CHARACTERISTIC	NOMINAL	REQUIREMENT	ACTUAL
Length	50 mm	± 0.0020 mm	49.9998 mm
Parallelism		0.0020 mm	0.0002 mm

Decision:

Actual measured results are within the requirements and accepted based on the below decision rule.

Decision rule basis:

Unless the actual results stated on the calibration certificate are highlighted with an "*" they will be classed as within the requirement / tolerance and accepted based on the binary decision rule (simple accept or simple reject) from JCGM 106: 2012 Clause 8.2 where the expanded uncertainty for $k = 2$ is less than or equal to the requirement / tolerance allowing zero guard bands. Any actual measured results highlighted with an "*" will be classed as outside the requirement / tolerance and rejected based on the above decision rule (see the attached document DOC-006-42.1). If the uncertainty of measurement is greater than the requirements / tolerance, no compliance statement will be provided.

Comments:


Traceable equipment used during calibration:

Identification Nr.	Description	Certificate Nr.
CZ 2	Metroscope	337172
WCL 120	Gauge Block Set	337194

** End of report **

JCGM 106: 2012 Clause 8.2 - Decision rule based on simple acceptance / shared risk

Unless the actual results stated on the calibration certificate are highlighted with an “*” or “A” they will be classed as within requirements / tolerance and **accepted** based on the binary decision rule (simple accept or simple reject) where the expanded uncertainty for $k = 2$ is less than or equal to the requirement / tolerance allowing zero guard bands. Actual measured results highlighted with an “*” or “A” are outside requirements / tolerance and **rejected** based on the above decision rule.



Based on JCGM 106: 2012

NOTE: If the uncertainty of measurement is greater than the required requirements / tolerance, no compliance statement will be provided. This will be clearly highlighted on the calibration certificate.