

CERTIFICATE OF CALIBRATION

Issue:- Certificate Number: **99015**
99015_10 Date of Issue: **01-Sep-26**
Approved Signatory: **Thomas Herrington**
Page 1 of 2 Signed: 



Submitter:-

Mecmesin Limited
Newton House
Spring Copse Business Park
Slinfold
West Sussex
RH13 0SZ

Issued by:-

Kent Scientific Services
8 Abbey Wood Road
Kings Hill
West Malling
Kent
ME19 4YT
Tel: 03000 415 100
Fax: 01732 220006

EQUIPMENT: Weights
SERIAL NUMBER: ~~AH8~~ Set AH1
MAKE/TYPE: N/A
STANDARDS USED: Set 12412
DATE RECEIVED: 26 August 2026
DATE CALIBRATED: 1 September 2026
DETAILS: 14 Cast Iron, 1 Brass

MEASUREMENTS:

Kent Scientific Services method used: CAL SMALL, Calibration of Small Masses.

The calibrations took place in a controlled environment with the temperature held between 18°C and 22°C, and with the relative humidity held between 40% and 60%.

The measurement results obtained in the table, where each measured value given represents not the true mass, but the mass of a hypothetical weight of density 8,000 kg.m⁻³, which in air of density 1.2 kg.m⁻³ would balance the corresponding weight identified in the first column at 20°C.

The method of weighing was by substitution (Borda's method). In each instance the standard weight used had been calibrated by UKAS Calibration Laboratory number 0474, 0260 or 0352 within the previous three years. The uncertainty of measurements for each of the different denominations is listed in the last column of the table. Duplicate weights, where present, are indicated by a dot or dots.

Customer supplied information is notated with a ~, and results relate only to the item(s) calibrated.

Unless otherwise notated, samples are tested in as received condition at Kent Scientific Services.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

TABLE OF MEASUREMENT RESULTS

Identity Mark	Nominal Mass	Measured Value	Error from Nominal	Estimated Uncertainty
M15	5N	509.648 4 g	+ 40.2 mg	± 5.1 mg
M14	100N	10192.12 g	- 40 mg	± 100 mg
M13	100N	10191.94 g	- 220 mg	± 100 mg
M12	100N	10192.00 g	- 160 mg	± 100 mg
M11	100N	10192.23 g	+ 60 mg	± 100 mg
M10	50N	5096.076 g	- 5 mg	± 52 mg
M09	20N	2038.311 g	- 122 mg	± 20 mg
M08	20N	2038.379 g	- 54 mg	± 20 mg
M06	10N	1019.140 g	- 76 mg	± 10 mg
M07	5N	509.541 7 g	- 66.5 mg	± 5.1 mg
M05	1N	101.918 7 g	- 3.0 mg	± 1.0 mg
M03	1N	101.920 2 g	- 1.4 mg	± 1.0 mg
M02	1N	101.909 4 g	- 12.2 mg	± 1.0 mg
S2E	1N	101.920 3 g	- 1.3 mg	± 1.0 mg
MB4	0.5N	50.902 10 g	- 58.72 mg	± 1.61 mg
M09 *	20N	2038.445 g	+ 12 mg	± 20 mg
M06 *	10N	1019.213 g	- 4 mg	± 10 mg
M07 *	5N	509.599 2 g	- 8.9 mg	± 5.1 mg
M02 *	1N	101.919 4 g	- 2.2 mg	± 1.0 mg
M15 *	5N	509.623 4 g	+ 15.2 mg	± 5.1 mg

* Denotes post adjustment calibration

The basis for conversion between force units and mass units is that a 1kg mass will experience a force of g newtons where g is the strength of the local gravitational field. At Kent Scientific Services the estimated local $g = 9.81146 \text{ ms}^{-2}$.

END OF RESULTS