CERTIFICATE OF CALIBRATION

Issue:-	Certificate Number:	**99056**
99056_10	Date of Issue:	**15-Sep-26**
	Approved Signatory:	**Thomas Herrington**
Page 1 of 2	Signed:	

UKAS CALIBRATION
0352

Kent County Council
kent.gov.uk

Submitter:-
Mecmesin Limited
Newton House
Spring Copse Business Park
Slinfold
West Sussex
RH13 0SZ

Issued by:-
Kent Scientific Services
8 Abbey Wood Road
Kings Hill
West Malling
Kent
ME19 4YT
Tel: 03000 415 100
Fax: 01732 220006

EQUIPMENT: Weights

SERIAL NUMBER: Set AH9

MAKE/TYPE: N/A

STANDARDS USED: Set 12412

DATE RECEIVED: 10 September 2026

DATE CALIBRATED: 15 September 2026

DETAILS: 10 Brass

MEASUREMENTS:

Kent Scientific Services method used: CAL SMALL, Calibration of Small Masses.
The calibrations took place in a controlled environment with the temperature held between 18°C and 22°C, and with the relative humidity held between 40% and 60%.
The measurement results obtained in the table, where each measured value given represents not the true mass, but the mass of a hypothetical weight of density 8,000 kg.m^{-3}, which in air of density 1.2 kg.m^{-3} would balance the corresponding weight identified in the first column at 20°C.
The method of weighing was by substitution (Borda's method). In each instance the standard weight used had been calibrated by UKAS Calibration Laboratory number 0474, 0260 or 0352 within the previous three years.
The uncertainty of measurements for each of the different denominations is listed in the last column of the table.
Duplicate weights, where present, are indicated by a dot or dots.
Customer supplied information is notated with a ~, and results relate only to the item(s) calibrated.
Unless otherwise notated, samples are tested in as received condition at Kent Scientific Services.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

TABLE OF MEASUREMENT RESULTS

Nominal Mass	Measured Value	Error from Nominal	Estimated Uncertainty
100 g	99.998 3 g	- 1.7 mg	± 1.0 mg
100 g°	100.000 5 g	+ 0.5 mg	± 1.0 mg
50 g	49.999 73 g	- 0.27 mg	± 0.60 mg
20 g	20.000 67 g	+ 0.67 mg	± 0.50 mg
20 g°	20.000 61 g	+ 0.61 mg	± 0.50 mg
10 g	10.000 49 g	+ 0.49 mg	± 0.40 mg
5 g	5.000 89 g	+ 0.89 mg	± 0.30 mg
2 g	2.000 34 g	+ 0.34 mg	± 0.24 mg
2 g°	2.000 51 g	+ 0.51 mg	± 0.24 mg
1 g	1.000 31 g	+ 0.31 mg	± 0.20 mg

END OF RESULTS

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.