CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

DATE OF ISSUE: 28th October 2026 CERTIFICATE No: 1021278

UKAS CALIBRATION 0495

Lambda CALIBRATION LTD

11-13 Chorley Central
Business Park
Stump Lane
Chorley
PR6 0BL
Tel: 01257 244 670

APPROVED SIGNATORY

A.Scurr A.Smith M.Kenyon K.Quigley
H.Ramsay T.Ambrose M.Darbyshire

Customer: PPT Group UK Ltd, Halifax, HX3 6EP

Item No: 3075485 JH - 825

Description: A (used) Mitutoyo (167-133) External Micrometer Setting Piece (Flat Ended)

Date of Cal: 28/10/2026

Basis & Method: Lambda Procedure No: C.I.M-1A Pt.10

The setting piece was calibrated against standard units of length (gauge blocks / length bars).

Equipment Used: LMB-74-07, LMS-11-19, LMS-5-21

Temperature: 20°C ± 1°C

Visual Examination/Suitability for Calibration

Instrument Condition	Satisfactory

Measured Results (Deviation from Nominal Size)

Nominal Size	Deviation
825mm	-0.0004mm

Measured Results (Uniformity of Setting Disk or Parallelism of Flat Ended Setting Rod)

Nominal Size	Uniformity or Parallelism
825mm	0.0012mm

Results Summary: The reported results are as found.

Uncertainty of Measurement: Length: ±0.001mm + (0.005mm x L in M)
Parallelism: ±0.0002mm + (0.003mm x L in M)

End of Certificate

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. Unless otherwise stated: [1] The 'Compliance Statement' is based on 'simple acceptance' (result vs tolerance) with the relevant calibration uncertainty being no greater than the tolerance. [2] Reported activities were carried out at the address detailed in the header. [3] The results relate only to the items calibrated.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and / or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.